

WHAT TO CHECK

DEFIBRILLATORS SELF-TEST ON A DAILY, WEEKLY OR MONTHLY BASIS – SO YOUR DEVICE WILL SIGNAL, TYPICALLY WITH A FLASHING LIGHT OR AUDIBLE ALERT, IF THERE IS A PROBLEM, FOR EXAMPLE, PADS NOT CORRECTLY CONNECTED, OR LOW BATTERY.

HOW TO CHECK

SET UP A REGULAR INSPECTION PROCEDURE FOR YOUR DEFIBRILLATOR AND RECORD YOUR INSPECTIONS BELOW!

PADS	EXPIRY	DATE
------	--------	------

BATTERY EXPIRY DATE:

LOCATION:	
MODEL: —	
GUARDIAN:	
SERIAL NO:	

READY FOR USE?	ANY PROBLEMS DETECTED?	INITIALS	DATE
		ı	ı



NEED NEW PADS OR BATTERIES?

SCAN THE QR CODE TO VISIT **DEFIB PLANET** OR CALL **02922 402670** TO PURCHASE NEW ACCESSORIES BEFORE THEY EXPIRE!



CALON HEARTS UK
THE HEART SCREENING AND DEFIBRILLATOR
CHARITY
CHARITY MARKER - 1193404